

Certificate of Analysis

4X Bolt™ LDS Sample Buffer

Product No. 15764202, B0007, 13276499

Lot No. 2933076

Date of Manufacture 20-Jun-2024

Expiration Date 30-Jun-2025

pH of (4X) Solution:

Specification: (8.41 - 8.53)

Result: Meets specification.

Refractive Index of (4X) Solution

Specification: (49.8 - 51.0) %BRIX

Result: Meets specification.

Silver Staining:

Specification: Exhibits no contaminant bands upon silver staining.

Result: Meets specification

Overall Result:

PASS

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 01-Jul-2024