

Certificate of Analysis

4X Bolt™ LDS Sample Buffer

Product No. 15774202, B0008, 13286499

Lot No. 2896343

Date of Manufacture 11-Jul-2024

Expiration Date 31-Jul-2025

pH of (4X) Solution:

Specification: (8.41 - 8.53)

Result: Meets Specification

Silver Staining:

Specification: Exhibits no contaminant bands upon silver staining.

Result: Meets Specification

Overall Result:

PASS

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevoyn Joseph
Director, Quality
Issued on 17-Jul-2024