

Certificate of Analysis

Taq DNA Polymerase, native

Product No. 18038018, 10388482

Lot No. 3018317

Date of Manufacture 09-Sep-2024

Expiration Date 09-Sep-2026

PCR Functional Assay

Specification: Negative control lanes must be free of visible product and an 800 bp band must be visible in at least one lane of each target amount tested.

Result: Meets specifications.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevoyn Joseph
Director, Quality
Issued on 12-Sep-2024