

# Certificate of Analysis

## M-MLV Reverse Transcriptase Buffer Set

Product No. 18057018, 10512703

Lot No. 2969300

Date of Manufacture 13-May-2024

---

### Warning:

This product is distributed for laboratory research use only. CAUTION: Not for diagnostic use. The safety and efficacy of this product in diagnostic or other clinical uses has not been established.

### Performance in unit assay

Specification: When the 5X First Strand Buffer and 0.1 M DTT are used in an M-MLV Reverse Transcriptase unit assay with a qualified lot of enzyme, the unit specification for the enzyme must be met.

Results: Meets specification

### Ribonuclease Assay

Specification: When the 5X First Strand Buffer and 0.1 M DTT are incubated with 5 µg of 0.24-9.5 Kb RNA ladder at 37°C for 30 minutes, there must be no downward smearing below the 0.24-kb band as compared to the controls.

Results: Meets specification

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevoyn Joseph  
Director, Quality  
Issued on 15-May-2024