

Certificate of Analysis

Random Primers DNA Labeling System

Product No. 18187013, 18187-013, 10337872

Lot No. 2484463

Date of Manufacture 16-Mar-2022

Incorporation Assay

Specification: Using [$\alpha^{32}\text{P}$]dCTP in a random primer reaction, the system must yield $\geq 23\%$ incorporation (equivalent to $\geq 1.02 \times 10^9$ cpm/ μg) within 1 hour.

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevoyn Joseph
Director, Quality
Issued on 29-Jan-2025