

Certificate of Analysis

Terrific Broth

Product No. 22711022, 11538916

Lot No. 2306AU

Date of Manufacture 28-Jun-2023

DCM Appearance

Specification: Homogeneous, free flowing, beige to yellow-beige

Result: Meets specification.

PPM Appearance

Specification: Clear, yellow to amber, no precipitate

Result: Meets specification.

pH of 4.76% solution at 25° C

Specification: 7.2 +/- 0.2

Result: Meets specification.

Microorganism Growth

Specification: *Escherichia coli* ATCC 23724, good to excellent growth at incubation temp of 35° C

Result: Meets specification.

Specification: *Escherichia coli* ATCC 33694, good to excellent growth at incubation temp of 35° C

Result: Meets specification.

Specification: *Escherichia coli* ATCC 33849, good to excellent growth at incubation temp of 35° C

Result: Meets specification.

Specification: *Escherichia coli* ATCC 39403, good to excellent growth at incubation temp of 35° C

Result: Meets specification.

Specification: *Escherichia coli* ATCC 47014, good to excellent growth at incubation temp of 35° C

Result: Meets specification.

Specification: *Escherichia coli* ATCC 53868, good to excellent growth at incubation temp of 35° C

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 05-Jan-2024