

Certificate of Analysis

RiboMinus™ Eukaryote Kit for RNA-Seq

Product No. A1083708, 10063854

Lot No. 2570487

Date of Manufacture 19-Dec-2022

The RiboMinus™ Eukaryote Kit for RNA-Seq is functionally tested for optimum performance using the following tests:

Total RNA Assay

Analysis of the molecular size distribution of RNA from depleted samples must show 18S and 28S ribosomal RNA depletion of 95% (\pm 5%) as compared to non-depleted samples.

Results: Meets specification.

Gel Electrophoresis

The depleted RNA sample is analyzed by gel electrophoresis. By visual inspection, no 18S and 28S ribosomal RNA must be present in the depleted samples.

Results: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 29-Jan-2025