

Certificate of Analysis

PureLink™ Fast Low-Endotoxin Midi Plasmid Purification Kit (25 preps)

Product No. A35892, 15900493

Lot No. 24026923

Date of Manufacture 02-Oct-2024

Expiration Date 01-Oct-2026

Midiprep Column Assemblies

Method: PureLink™ Fast Low-Endotoxin Kits qualified by isolating a high copy number control plasmid DNA from 50ml bacterial culture grown overnight in LB media

Specification: A260/280 ratio > 1.8.

Result: Meets Specification

Specification: A260/230 ratio > 1.8.

Result: Meets Specification

Specification: > 145 µg recovered plasmid (at > 725 ng/µl).

Result: Meets Specification

Specification: Column Assemblies are undamaged and no foreign material present within the column assemblies

Result: Meets Specification

Buffers/Reagents

Method: PureLink™ Fast Low-Endotoxin Kits qualified by isolating a high copy number control plasmid DNA from 50ml bacterial culture grown overnight in LB media

Specification: A260/280 ratio \geq 1.8.

Result: Meets Specification

Specification: A260/230 ratio \geq 1.8.

Result: Meets Specification

Specification: \geq 145 µg recovered plasmid (at \geq 725 ng/µl).

Result: Meets Specification

Specification: No visible debris present in any cleared lysate.

Result: Meets Specification

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com

A handwritten signature in blue ink, appearing to read 'Chevohn Joseph', with a long horizontal flourish extending to the right.

Chevohn Joseph
Director, Quality
Issued on 15-Oct-2024