

# Certificate of Analysis

## No-Stain™ Protein Labeling Reagent (Trial Size)

Product No. A44717, 16351358

Lot No. 2808782

Date of Manufacture 24-Oct-2023

---

### Linearity

Specification: Following quantification of a dilution series of protein in a gel transferred to a membrane and labeled using the No-Stain™ Protein Labeling Reagent, analysis shall yield an  $R^2$  value  $\geq 0.90$  for linear fit of  $\geq 4$  data points.

Result: Meets specification.

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevohn Joseph  
Director, Quality  
Issued on 02-Nov-2023