

# Certificate of Analysis

## 20X Bolt™ MOPS SDS Running Buffer

Product No. B000102, 13236499, 15774462

Lot No. 2960908

Date of Manufacture 17-Jul-2024

Expiration Date 17-Jul-2025

---

### pH of 20X Solution:

Specification: ( 7.67 - 7.77 )

Result: Meets Specification

### pH of 1X Solution:

Specification: ( 7.54 - 7.64 )

Result: Meets Specification

### Refractive Index of 20X solution:

Specification: ( 31.3 - 32.3 )% Brix

Result: Meets Specification

### Overall Result:

PASS

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevohn Joseph  
Director, Quality  
Issued on 25-Jul-2024