

# Certificate of Analysis

## 20X Bolt™ MOPS SDS Running Buffer

Product No. B0001, 13226499, 15734202

Lot No. 2966954

Date of Manufacture 05-Aug-2024

Expiration Date 06-Aug-2025

---

### pH of 20X Solution:

Specification: ( 7.67 - 7.77 )

Result: Meets specification.

### pH of 1X Solution:

Specification: ( 7.54 - 7.64 )

Result: Meets specification.

### Refractive Index:

Specification: ( 31.3 - 32.3 ) % Brix

Result: Meets specification.

### Overall Result:

PASS

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevohn Joseph  
Director, Quality  
Issued on 14-Aug-2024