

Certificate of Analysis

4X Bolt™ LDS Sample Buffer

Product No. B0007, 13276499, 15764202

Lot No. 2820664

Date of Manufacture 20-Dec-2023

Expiration Date 31-Dec-2024

pH of (4X) Solution:

Specification: (8.41 - 8.53)

Result: Meets specification.

Refractive Index of (4X) Solution

Specification: (49.8 - 51.0) %BRIX

Result: Meets specification.

Silver Staining:

Specification: Exhibits no contaminant bands upon silver staining.

Result: Meets specification

Overall Result:

PASS

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 29-Jan-2024