

Certificate of Analysis

PureLink™ Genomic Digestion Buffer

Product No. K182301, 10468843

Lot No. 2579991

Date of Manufacture 30-Jan-2023

The PureLink™ Genomic Digestion Buffer is functionally tested for optimum performance using the following tests:

Gel Electrophoresis

DNA is analyzed by gel electrophoresis. A single band of DNA must be present and must be larger than 12 Kb compared to the 1 Kb Plus DNA Ladder.

Result: Meets Specification

A_{260/280} Ratio

Must be between 1.7-2.2.

Result: Meets Specification.

Yield

>4 µg of genomic DNA must be obtained from 1 x 10⁹ Top10 E.coli cells.

Result: Meets Specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 29-Jan-2025