

Certificate of Analysis

PureLink™ Pro 96 Plasmid Purification Kit

Product No. K211004A, 10063814

Lot No. 2792870

Date of Manufacture 27-Nov-2023

The PureLink™ Pro 96 Plasmid Purification Kit is functionally tested for optimum performance using the following tests:

A₂₆₀/A₂₈₀ Ratio

Must be greater than or equal to 1.8

Result: meets specification

Gel Electrophoresis

Genomic DNA and RNA should not be visible in test samples.

Purified plasmid samples should be at least 60% supercoil DNA with no noticeable smearing or shift in form upon incubation at 37° C in the presence of MgCl₂

Result: Meets Specification

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevoyn Joseph
Director, Quality
Issued on 05-Dec-2023