

Certificate of Analysis

PureLink™ PCR Micro Kit

Product No. K310250, 10043063

Lot No. 2918356

Date of Manufacture 15-Jul-2024

The PureLink™ PCR Micro Kit is functionally tested for optimum performance using the following tests:

DNA Recovery

Binding Buffer: Greater than or equal to 80% of the mass of a Low DNA Mass Ladder loaded onto the column must be recovered post-purification.

Results: Meets specification.

Gel Electrophoretic Analysis

Low DNA Mass Ladder fragments and primers ranging in size from 33 bases to 2,000 base pairs are bound to and eluted from columns following the instructions in the product manual.

Binding Buffer: No 33 base primers must be visible in samples bound with Binding Buffer.

Results: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 31-Jul-2024