

Certificate of Analysis

Nitrocellulose/Filter Paper Sandwich, 0.2 µm, 8.3 x 7.3 cm

Product No. LC2000, 10368292

Lot No. 2982994

Date of Manufacture 25-Jul-2024

Dimensional Requirements

Specifications: (8.15 - 8.45 cm) x (7.15 - 7.50 cm)

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 26-Jul-2024