

Certificate of Analysis

MagicMark™ XP Western Protein Standard

Product No. LC5603, 10610856

Lot No. 2893793

Date of Manufacture 21-Mar-2024

Expiration Date 16-Nov-2024

Gel Testing

Specification: Nine protein bands must be present in a 5µl load when tested on 10% Bis-Tris gel with MOPS running buffer and stained with Simply Blue SafeStain.

Result: Meets specification.

Western Blot Test

Specification: Nine protein bands must be present in a 2µl load after Western blotting and detection with the WesternBreeze Anti-Mouse Chromatic or Chemiluminescent Kit.

Result: Meets specification.

Intensity

Specification: Band intensity from the lot being tested on the 2 µl load must be comparable to the protein bands from previously approved lot (in the 30 second film in the Chemiluminescent Test).

Result: Meets specification.

Contamination

Specification: Contaminating band intensity must be less than 20% of major band intensity.

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com

A handwritten signature in blue ink, appearing to read 'Chevohn Joseph', with a long horizontal flourish extending to the right.

Chevohn Joseph
Director, Quality
Issued on 02-Apr-2024