

Certificate of Analysis

SeeBlue™ Pre-stained Protein Standard

Product No. LC5625, 10237352

Lot No. 2968646

Date of Manufacture 26-Jun-2024

Expiration Date 12-Apr-2025

Gel Testing

10%BT, 12%TG, and 10-20%TR.

Nine bands must be present on the 10%BT and 10-20%TR.

Gel test performance must be comparable to 2 previously released lots.

Intensity

Band intensity must not differ more than 20% from 2 previously released lots.

Migration

Bands must migrate to within one half band width of the reference lots.

Contamination

Contaminant bands must be less than 20% of the intensity of the major bands, except for those associated with insulin.

For insulin A, contaminant band must be less than 25% of the intensity of the major bands in previously released lots.

Results: Meets all specifications.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com

A handwritten signature in blue ink, appearing to read 'Chevohn Joseph', with a long horizontal flourish extending to the right.

Chevohn Joseph
Director, Quality
Issued on 27-Jun-2024