

# Certificate of Analysis

## Mark12™ Unstained Standard

Product No. LC5677, 10154962

Lot No. 2893774

Date of Manufacture 15-Apr-2024

Expiration Date 03-Feb-2025

---

### Gel Testing

8% TG, 10-20% TR, and 10% BT.

Twelve bands must be present on the 10-20% TR and the 10% BT.

Gel test performance must be comparable to 2 previously released lots.

### Intensity

Band intensity when stained must not differ more than 20% from 2 previously released lots.

### Migration

Bands must be straight and sharp.

### Contamination

Contaminant bands must be less than 20% of the intensity of the major bands.

### Overall Testing

Pass

This certifies that the Mark12 Standard meets all QC specifications as stated above.

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevohn Joseph  
Director, Quality  
Issued on 15-Apr-2024