

Certificate of Analysis

Mark12™ Unstained Standard

Product No. LC5677, 10154962

Lot No. 3019108

Date of Manufacture 18-Sep-2024

Expiration Date 28-Apr-2025

Gel Testing

8% TG, 10-20% TR, and 10% BT.

Twelve bands must be present on the 10-20% TR and the 10% BT.

Gel test performance must be comparable to 2 previously released lots.

Intensity

Band intensity when stained must not differ more than 20% from 2 previously released lots.

Migration

Bands must be straight and sharp.

Contamination

Contaminant bands must be less than 20% of the intensity of the major bands.

Overall Testing

Pass

This certifies that the Mark12 Standard meets all QC specifications as stated above.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 19-Sep-2024