

Certificate of Analysis

HiMark™ Pre-stained Protein Standard

Product No. LC5699, 10237532

Lot No. 2980919

Date of Manufacture 16-Sep-2024

Expiration Date 19-May-2025

Gel Testing

Specification: Nine bands must be present on both a 7% and 3-8% Tris-Acetate gels. Performance must be comparable to two previously released lots.

Results: Meet specifications.

Migration

Specification: Band migration must be straight and sharp, and within 1/2 bandwidth of two previously released lots.

Results: Meet specifications.

Intensity

Specification: Band intensity must not differ more than 20% from two previously released lots.

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 19-Sep-2024