

Certificate of Analysis

Novex™ Sharp Pre-stained Protein Standard

Product No. LC5800, 10640504

Lot No. 2750848

Date of Manufacture 19-Jun-2024

Expiration Date 19-May-2025

Gel Analysis

Specifications: There must be 12 protein bands present on a NuPAGE® Novex 4-12% Bis-Tris Gel run with 1X MES buffer. There must be at least 11 bands present on 4-12% Bis-Tris Gel run with 1X MOPS buffer and 4-20% Tris-Glycine Gel run with 1X Tris-Glycine SDS buffer.

Results: Meet specifications.

Specification: Any contaminant band must be less than 10% in intensity to the nearest major band.

Result: Meets specification.

Specification: Prestained protein band from lot being tested must migrate within 1/2 bandwidth from prior lot.

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 20-Jun-2024