

# Certificate of Analysis

Empty Gel Cassettes, mini, 1.5 mm

Product No. NC2015, 11559156, 361041737

Lot No. 2867172

Date of Manufacture 15-Feb-2024

Expiration Date 15-Feb-2025

---

## Testing Method

Samples are assessed for ease of opening and seal integrity, after joining.

Samples are assessed for volume and comb gap.

## Specifications

Cassettes must open without shattering.

Cassettes must not leak.

Cassette must accomodate 10.0 mL - 11.75 mL of fluid

Comb gap in cassette must be 1.425 mm - 1.575 mm

## Overall Result

Meets Specification

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevohn Joseph  
Director, Quality  
Issued on 19-Feb-2024