

Certificate of Analysis

Empty Gel Cassettes, mini, 1.5 mm

Product No. NC2015, 11559156, 361041737

Lot No. 2812820

Date of Manufacture 18-Dec-2023

Expiration Date 13-Mar-2025

Testing Method

Samples are assessed for ease of opening and seal integrity, after joining.

Samples are assessed for volume and comb gap.

Specifications

Cassettes must open without shattering.

Cassettes must not leak.

Cassette must accomodate 10.0 mL - 11.75 mL of fluid

Comb gap in cassette must be 1.425 mm - 1.575 mm

Overall Result

Meets Specification

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevoyn Joseph
Director, Quality
Issued on 23-Mar-2024