

# Certificate of Analysis

NuPAGE® MOPS SDS Running Buffer (20X), 500 ml

Product No. NP0001, 11589156, 15724302

Lot No. 2966953

Date of Manufacture 05-Aug-2024

Expiration Date 06-Aug-2025

---

**pH of 20X Solution:**

Specification: ( 7.67 - 7.77 )

Results: Meets specification.

**pH of 1X Solution:**

Specification: ( 7.54 - 7.64 )

Results: Meets specification.

**Refractive Index:**

Specification: ( 31.3 - 32.3 ) % Brix

Results: Meets specification.

**Overall Results:**

PASS

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevohn Joseph  
Director, Quality  
Issued on 14-Aug-2024