

Certificate of Analysis

Bolt™ Empty Mini Gel Cassettes

Product No. NW2010, 15394604

Lot No. 2862965

Date of Manufacture 21-Mar-2024

Testing Method

Samples are assessed for ease of opening and seal integrity, after joining.

Samples are assessed for volume and comb gap.

Specifications

Cassettes must open without shattering.

Cassettes must not leak.

Cassette must accomodate 7.90 mL - 8.50 mL of fluid

Comb gap in cassette must be 0.925 mm - 1.075 mm

Overall Result

Meets Specification

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevoyn Joseph
Director, Quality
Issued on 21-Mar-2024