

Certificate of Analysis

WesternBreeze® Chemiluminescent Immunodetection Kit (anti-Rabbit)

Product No. WB7106, 10542703

Lot No. 2926431

Date of Manufacture 12-Apr-2024

Expiration Date 01-Oct-2024

Specification:

- * Solutions meet in-process specifications
- * Non-specific binding must be less intense than the 100 pg band of Human IgG for Rabbit Kit
- * Background comparable between test and control for intensity and evenness
- * 100 pg load of Human IgG detectable for Rabbit Kit in < 30 minutes with Chemiluminescence.

PASS

Overall Result:

PASS

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevoyn Joseph
Director, Quality
Issued on 15-Apr-2024