

Certificate of Analysis

WesternBreeze® Chemiluminescent Immunodetection Kit (anti-Rabbit)

Product No. WB7106, 10542703

Lot No. 2983013

Date of Manufacture 13-Aug-2024

Expiration Date 17-Jun-2025

Specification: * Solutions meet in-process specifications* Non-specific binding must be less intense than the 100 pg band of Human IgG for Rabbit Kit * Background comparable between test and control for intensity and evenness * 100 pg load of Human IgG detectable for Rabbit Kit in < 30 minutes with Chemiluminescence.PASS

Overall Result: PASS

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevohn Joseph
Director, Quality
Issued on 13-Aug-2024