

Certificate of Analysis

Empty Gel Cassettes with Combs, midi, 12 + 2 well

Product No. WC1012, 12847582

Lot No. 25010600

Date of Manufacture 06-Jan-2025

Expiration Date 06-Jan-2026

Testing Method

Samples are assessed for ease of opening and seal integrity, after joining.

Samples are assessed for volume and comb gap.

Specifications

Cassettes must open without shattering.

Cassettes must not leak.

Cassette must accomodate 11.50 mL - 13.20 mL of fluid

Comb gap in cassette must be 0.925 mm - 1.075 mm

Overall Result

Meets Spefication

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Chevoyn Joseph
Director, Quality
Issued on 30-Jan-2025