

# Certificate of Analysis

Empty Gel Cassettes with Combs, midi, 12 + 2 well

Product No. WC1012, 12847582

Lot No. 23042602

Date of Manufacture 26-Apr-2023

Expiration Date 25-Apr-2024

---

## **Testing Method**

Samples are assessed for ease of opening and seal integrity, after joining.

Samples are assessed for volume and comb gap.

## **Specifications**

Cassettes must open without shattering.

Cassettes must not leak.

Cassette must accomodate 11.50 mL - 13.20 mL of fluid

Comb gap in cassette must be 0.925 mm - 1.075 mm

## **Overall Result**

Meets Spedefication

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevoyn Joseph  
Director, Quality  
Issued on 29-Jan-2025