

Certificate of Analysis

Product Name: Dimension Chemistry II Calibrator DC20
Product Number: DC20
Lot Number: 5DN084

Expiration Date: 10/01/2026
Storage: 2 - 8°C

Description	Lot Number	Expiration Date
DC20 Chem II Level 1	5DN184	2026-10-01
DC20 Chem II Level 2	5DN284	2026-10-01
DC20 Chem II Level 3	5DN384	2026-10-01

Specification	Specification Limit	Result
Level 1: Magnesium	0.0 - 0.3 mg/dL	0.0
Level 1: Phosphorus	1.95 - 2.05 mg/dL	2.00
Level 1: Triglycerides	117.5 - 122.5 mg/dL	120.0
Level 1: pH @ 25 °C	1.4 - 1.8	1.6
Level 2: Magnesium	8.85 - 9.15 mg/dL	9.00
Level 2: Phosphorus	4.95 - 5.05 mg/dL	5.00
Level 2: Triglycerides	237.5 - 242.5 mg/dL	240.0
Level 2: pH @ 25 °C	1.4 - 1.8	1.7
Level 3: Magnesium	17.80 - 18.20 mg/dL	18.00
Level 3: Phosphorus	7.90 - 8.10 mg/dL	8.00
Level 3: Triglycerides	480.0 - 490.0 mg/dL	485.0
Level 3: pH @ 25 °C	1.4 - 1.8	1.7

Customer Name: Siemens Healthcare Diagnostics

Siemens Level Part Number: DC20 (792020.955)

The Siemens Level lot number is the same lot number shown above in the upper left portion of the certificate.

Beth Hochman

Quality Representative

Date: 2025-05-28