

Certificate of Analysis

Product Name: Anti-Tg Microparticle Reagent
OEM Part Number: 106523
OEM Lot Number: 42699UN25
Storage: 2 - 8°C

Abbott List Number: 106523
Abbott Commodity Number: 2K46G
Expiration Date: 2025-06-16
Package Insert No./Rev. N/A / N/A
Test Date 2025-04-21


Specification	Specification Limit	Result
Calibrator B/A Ratio	≥ 10.0	23.8 Pass
Calibrator F/A Ratio	≥ 1283	2892 Pass
Negative Control	Lower Limit	0.19 IU/mL
Negative Control	Upper Limit	0.99 IU/mL
Negative Control	Result	0.77 IU/mL Pass
Negative Control	Lot Number	759932
Positive Control	Lower Limit	110 IU/mL
Positive Control	Upper Limit	186 IU/mL
Positive Control	Result	148 IU/mL Pass
Positive Control	Lot Number	759933
Assay Diluent	Lot Number	42761UN25
Conjugate	Lot number	42700UN25
Quantity Manufactured	Report	79.150 L
Date of Manufacture	Report	2025-04-17

For in-vitro diagnostic use.

Expiration Date above is the Fill by Date of bulk.

Anti-Tg Reagents contain animal and human source material.

This lot meets all current release specifications per quality document MS.MASTER342. The above product has been manufactured in accordance with Fisher Diagnostics' operating procedure.



Quality Representative

Date: 2025-04-24