

## *Certificate of Analysis*

Product Name: Anti-Tg Microparticle Reagent  
OEM Part Number: 106523  
OEM Lot Number: 57277UN25  
Storage: 2 - 8°C

Abbott List Number: 106523  
Abbott Commodity Number: 2K46G  
Expiration Date: 2025-07-28  
Package Insert No./Rev. N/A / N/A  
Test Date 2025-06-23


Specification	Specification Limit	Result
Calibrator B/A Ratio	$\geq 10.0$	19.1 Pass
Calibrator F/A Ratio	$\geq 1283$	2323 Pass
Negative Control	Lower Limit	0.19 IU/mL
Negative Control	Upper Limit	0.99 IU/mL
Negative Control	Result	0.82 IU/mL Pass
Negative Control	Lot Number	759932
Positive Control	Lower Limit	110 IU/mL
Positive Control	Upper Limit	186 IU/mL
Positive Control	Result	150 IU/mL Pass
Positive Control	Lot Number	759933
Assay Diluent	Lot Number	57279UN25
Conjugate	Lot number	57278UN25
Quantity Manufactured	Report	67.750 L
Date of Manufacture	Report	2025-05-29

For in-vitro diagnostic use.

Expiration Date above is the Fill by Date of bulk.

Anti-Tg Reagents contain animal and human source material.

This lot meets all current release specifications per quality document MS.MASTER342. The above product has been manufactured in accordance with Fisher Diagnostics' operating procedure.



Quality Representative

Date: 2025-06-27