

CERTIFICATE OF ANALYSIS
DRI Hydrocodone 3 X 18 mL EIA Kit

Catalog #	10018054
Kit Lot #	74758982
Expiration Date (YYYY-MM-DD)	2024-04-30

Performance Testing

225 ng/mL Control
375 ng/mL Control

Result	Specifications (ng/mL)
248 ng/mL	180-270
394 ng/mL	300-450


Quality Control

11/29/2022
Date

10018054COA REV 2
Effective Date: 09/26/16

Legal Entity: Microgenics Corporation

Thermo Fisher Scientific
46500 Kato Road
Fremont, CA 94538 USA

Phone: 510-979-5000
Fax: 510-979-5002
Toll Free: 800-232-3342

Template: 10020092TMP, Rev. 1
Effective Date: 08/11/16