


**CERTIFICATE OF ANALYSIS**  
**DRI Hydrocodone 3 X 18 mL EIA Kit**

|                              |            |
|------------------------------|------------|
| Catalog #                    | 10018054   |
| Kit Lot #                    | 74848327   |
| Expiration Date (YYYY-MM-DD) | 2024-08-31 |

**Performance Testing**

225 ng/mL Control  
375 ng/mL Control

| Result    | Specifications (ng/mL) |
|-----------|------------------------|
| 247 ng/mL | 180-270                |
| 390 ng/mL | 300-450                |

  
Quality Control

03/27/2023  
Date

10018054COA REV 2  
Effective Date: 09/26/16

**Legal Entity: Microgenics Corporation**

Thermo Fisher Scientific  
46500 Kato Road  
Fremont, CA 94538 USA

Phone: 510-979-5000  
Fax: 510-979-5002  
Toll Free: 800-232-3342

Template: 10020092TMP, Rev. 1  
Effective Date: 08/11/16