

CERTIFICATE OF ANALYSIS
DRI Hydrocodone 3 X 18 mL EIA Kit

Catalog #	10018054
Kit Lot #	75110887
Expiration Date (YYYY-MM-DD)	2025-06-30

Performance Testing

225 ng/mL Control
375 ng/mL Control

Result	Specifications (ng/mL)
245 ng/mL	180-270
388 ng/mL	300-450


Quality Control

01/19/2024
Date