

Certificate of Analysis

Product Name:	AcroMetrix™ SARS-CoV-2 Control
Catalog Number:	954517
Lot Number:	042401
Date of Manufacture:	2024-04-18
Expiration Date:	2025-04-30
Analyte Source:	Heat Inactivated SARS-CoV-2
Number of Vials:	5 vials
Volume per Vial:	1.5 mL
Storage:	-20 °C or below

Summary of Release Testing Results

Extraction Method	Assay Method	Target Titer	Results
MagMAX™ Viral/Pathogen Nucleic Acid Isolation Kit	Digital PCR / In-House Assay	Positive	Pass

Precautions

Observe universal precautions for prevention of transmission of infectious agents when handling these materials.

Intended Use

This product is for Research Use Only (RUO), not for use in diagnostic procedures.

Biosafety Level

This material is BSL1.

CERTIFICATION

The product described above meets all performance specifications:

Completed By

Date

Approved By

Date