



Office 422 Aldo Ave., Santa Clara, CA 95054

Mfg. Site 3170 Molinaro St., Santa Clara, CA 95054

Product Certificate

Thermo Scientific
Fiberlite Products

This information is the product certification for item #

010-1491

Description:

Thermo Scientific Fiberlite Bottle 1000 mL PPCO (sets of 2)

Lot#:

0000885635

Issued Date

4/14/2025

Manufacture Date

4/14/2025

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
010-1407	Thermo Scientific Fiberlite Bottle 1000 mL	Component Part	-	-	-	
	Resin used for Bottle	Polypropylene copolymer	-	Passed	Passed	177.1520 (a)(3)(i) & (c)3.2(a) (use conditions A-H)
010-1471	Thermo Scientific Fiberlite Bottle 1000 mL Closure	Component Part	-	-	-	
	Resin used for closure	Nylon, 30% glass filled	-	Passed	Passed	177.1500
010-1462	Thermo Scientific Fiberlite Bottle 1000 mL Closure liner	Component Part	-	-	-	
	Resin used for closure liner	Polypropylene	-	Passed	Passed	177.1520(a)(3)(i) and (c)3.1a & 177.1520(b)
001-0298	Thermo Scientific Fiberlite Bottle 1000 mL O-ring	Component Part	-	-	-	
	Resin used for O-ring	Silicone	-	Passed	Passed	177.2600

Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity and USP Class VI.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol. If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol. The closure and plug and the silicone o-ring have been tested and shown to comply with USP Class VI requirements and to non-cytotoxic. The polypropylene/polycarbonate resin for the bottle meets the requirements of 21 CFR section 177.1520. No animal-derived components.