

GlutaMAX(TM) - I (100X)  
200mM

**Lot Number:** 2523109  
**Item Number:** 35050  
**Expiration Date:** 2025-03  
**Storage Temp:** 15 to 30C

For Research Use or Further Manufacturing. Not for diagnostic use or direct administration into humans or animals.

TEST	TEST ID	SPECIFICATION	RESULT	UNITS
<b>Osmolality</b>	OSMO0002	430 - 490	462	mOsm/kg
<b>pH</b>	PH0003	5.0 - 6.0	5.7	
<b>Sterility Testing</b>	STERI0007	Negative	Negative	

Read SDS

GIBCO brand, Thermo Fisher Scientific cell culture liquid products are prepared by an aseptic process for which each step has been validated to ensure that all products meet the industry standard sterility assurance level of  $10^{-3}$ ; i.e. product that demonstrates a contamination level of no more than 1 of 1,000 units during the manufacturing process. The highest level of sterility assurance (equal to or greater than  $10^{-6}$ ) cannot be achieved without terminal sterilization which is harmful to the performance of cell culture products.

*Clatz*

Quality Systems Department      Date: 30-Mar-2023

References

- OSMO0002: Thermo Fisher Scientific Specifications.
- PH0003: Thermo Fisher Scientific Specifications.
- STERI0007: Current Edition of USP, Thermo Fisher Scientific Modified.