



Certificate of Analysis

QC Code: GIBCO

TN Trypsin Neutralizer Solution
1X

Lot Number: 2539421
Item Number: R-002-100
Expiration Date: 2024-08
Storage Temp: -5 to -20C

For research use only. Not for use in diagnostic procedures.

TEST	TEST ID	SPECIFICATION	RESULT	UNITS
Appearance, Liquid - Clear, pink solution	SOLAPP0007	Acceptable	Acceptable	
Endotoxin Testing	ENDO0007	<=2.00	<0.10	EU/mL
Osmolality	OSMO0002	260 - 310	286	mOsm/kg
pH	PH0003	7.2 - 7.8	7.3	
Sterility Testing	STERI0007	Negative	Negative	
TE and TN Performance Test Growth	TETN0001	>=2.50	5.87	x 10 ⁴ cells/cm ²

Read SDS

A buffered, isotonic solution that contains 0.5% Newborn Bovine Serum. Suitable for neutralizing Trypsin/EDTA solution for the dissociation of cells from the culture surface.

Filtered via 0.2 um filtration.

EHotchkiss

Quality Systems Department

Date: 29-Sep-2022

References

- SOLAPP0007: Thermo Fisher Scientific Specifications.
- ENDO0007: Current United States Pharmacopeia, <85> Bacterial Endotoxins Test.
- OSMO0002: Thermo Fisher Scientific Specifications.
- PH0003: Thermo Fisher Scientific Specifications.
- STERI0007: Current Edition of USP, Thermo Fisher Scientific Modified.
- TETN0001: Cell culture tested.