

AF
Attachment Factor
1X

Lot Number: 2698772
Item Number: S-006-100
Expiration Date: 2025-05
Storage Temp: 2 to 8C

For research use only. Not for use in diagnostic procedures.

TEST	TEST ID	SPECIFICATION	RESULT	UNITS
Appearance, Liquid - Clear, colorless, proteinaceous solution	SOLAPP0007	Acceptable	Acceptable	
Endotoxin Testing	ENDO0007	<=5.00	2.45	EU/mL
Med 131, MVGS and AF Performance Test	MED1310001	>=2.50	3.32	x 10 ⁴ cells/cm ²
Sterility Testing	STERI0007	Negative	Negative	

Read SDS

An attachment factor for coating tissue culture dishes to enhance the growth of human microvascular endothelial cells.

Filtered via 0.2um filtration.

Hotchkiss

Quality Systems Department Date: 27-Jun-2023

References

- SOLAPP0007: Thermo Fisher Scientific Specifications.
- ENDO0007: Current United States Pharmacopeia, <85> Bacterial Endotoxins Test.
- MED1310001: Tested for growth performance using normal human microvascular endothelial cells.
- STERI0007: Current Edition of USP, Thermo Fisher Scientific Modified.