

# Certificate of Analysis

E-PAGE™ 8%, 3.7 mm, Midi Protein Gel, 48-well

Product No. EP04808, 10206172

Lot No. G290822-11

Date of Manufacture 29-Aug-2022

Expiration Date 29-May-2023

---

**Visual Test** (100% of cassettes)

Appearance

**Result**

Meets Specification

Electrical contacts

Passed test

**Functional Test** (sample testing)

run on E-Base™

Currents

Post staining\*

Resolution

Wells integrity

Protein sensitivity\*

Protein capacity\*

Straightness

\*With Coomassie® Brilliant Blue R-250 Dye according to Manual Protocol

**Results**

Normal range

Passed test

Passed test

Passed test

Min. 20 ng/band

Passed test

Passed test

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008

<https://www.thermofisher.com>

For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Nitzan Rozenblit

Sr. Manager, Manufacturing LT Israel

Issued on 29-Jan-2025