

Certificate of Analysis

E-PAGE™ 96 6%

Product No. EP09606, EPST9606, EPST96-06, 10654583

Lot No. P170322-11

Date of Manufacture 17-Mar-2022

Expiration Date 17-Sep-2022

<u>Visual Test</u> (100% of cassettes)	<u>Result</u>
Appearance	Meets Specification
Electrical contacts	Passed test

Functional Test
(sample testing) Results

<u>(run on E-Base™ EP)</u>	
Currents	Normal range
Post staining*	Passed test
Resolution	Passed test
Wells integrity	Passed test
Protein sensitivity*	Min. 20 ng/band
Protein capacity*	Passed test
Straightness	Passed test

*With Coomassie© R-250 according to Manual Protocol

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Nitzan Rozenblit
Sr. Manager, Manufacturing LT Israel
Issued on 29-Jan-2025