

Certificate of Analysis

E-PAGE™ 6%, 3.7 mm, Midi Protein Gel, 96-well

Product No. EP09606, 10654583

Lot No. P310822-01

Date of Manufacture 31-Aug-2022

Expiration Date 28-Feb-2023

Visual Test (100% of cassettes)

Appearance

Result

Meets Specification

Electrical contacts

Passed test

Functional Test (sample testing)

run on E-Base™

Currents

Normal range

Post staining*

Passed test

Resolution

Passed test

Wells integrity

Passed test

Protein sensitivity*

Min. 20 ng/band

Protein capacity*

Passed test

Straightness

Passed test

*With Coomassie® Brilliant Blue R-250 Dye according to Manual Protocol

Results

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008

<https://www.thermofisher.com>

For inquiries, contact us at cofarequests@thermofisher.com



Nitzan Rozenblit

Sr. Manager, Manufacturing LT Israel

Issued on 29-Jan-2025