

Certificate of Analysis

E-Gel™ EX Agarose Gels, 1%

Product No. G401001, 11571186

Lot No. T121224-01

Date of Manufacture 12-Dec-2024

Expiration Date 12-Dec-2025

Visual Test (100% of cassettes)

Appearance

Result

Meets Specification

Electrical contacts

Passed test

Functional Test (sample testing) run on PowerSnap™

Currents

Normal range

SYBR® Gold background

Passed test

Resolution

Passed test

Well integrity

Passed test

DNA sensitivity

Min. 0.5 ng/band

DNA capacity

Passed test

Uniformity of running front

Passed test

Opening

Passed test

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Nitzan Rozenblit
Sr. Manager, Manufacturing LT Israel
Issued on 30-Jan-2025