

Certificate of Analysis

E-Gel™ CloneWell™ II Agarose Gels with SYBR Safe, 0.8%

Product No. G661818, 15796548

Lot No. 2W060125-02

Date of Manufacture 06-Jan-2025

Expiration Date 06-Oct-2025

Visual Test (100% of cassettes)

Appearance

Electrical contacts

Result

Meets Specification

Passed test

Functional Test (sample testing)

run on PowerSnap™

Currents

SYBR® Safe background

Resolution

Well integrity

DNA sensitivity

DNA capacity

Uniformity of running front

Time to reference line

Results

Normal range

Passed test

Passed test

Passed test

Min. 5 ng/band

Passed test

Passed test

Within range

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific

Life Sciences Solutions

5781 Van Allen Way

Carlsbad, CA, USA 92008

<https://www.thermofisher.com>

For inquiries, contact us at cofarequests@thermofisher.com



Nitzan Rozenblit

Sr. Manager, Manufacturing LT Israel

Issued on 30-Jan-2025