

Certificate of Analysis

E-Gel® CloneWell 0.8% II with SYBR™ Safe

Product No. G661818, 15796548

Lot No. 2W200322-01

Date of Manufacture 20-Mar-2022

Expiration Date 20-Dec-2022

<u>Visual Test</u> _(100% of cassettes)	<u>Result</u>
Appearance	Meets Specification
Electrical contacts	Passed test

<u>Functional Test</u> _(sample testing) <u>run on E-GEL® iBase™</u>	<u>Results</u>
Currents	Normal range
Staining	Meets Specification
SYBR® Safe background	Passed test
Resolution	Passed test
Well integrity	Passed test
DNA sensitivity	Min. 5 ng/band
DNA capacity	Passed test
Uniformity of running front	Passed test
Time to reference line	Within range

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
<https://www.thermofisher.com>
For inquiries, contact us at cofarequests@thermofisher.com



Nitzan Rozenblit
Sr. Manager, Manufacturing LT Israel
Issued on 29-Jan-2025