

Certificate of Analysis

E-Gel™ 96 Gels with SYBR™ Safe DNA Gel Stain, 2%

Product No. G720802, 10063143

Lot No. N241224-01

Date of Manufacture 24-Dec-2024

Expiration Date 24-Sep-2025

Visual Test (100% of cassettes)

Appearance

Electrical contacts

Result

Meets Specification

Passed test

Functional Test (sample testing)

run on E-Base™

Currents

SYBR® Safe background

Resolution

Well integrity

DNA sensitivity

DNA capacity

Uniformity of running front

Results

Normal range

Passed test

Passed test

Passed test

Min. 5 ng/band

Passed test

Passed test

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008

<https://www.thermofisher.com>

For inquiries, contact us at cofarequests@thermofisher.com



Nitzan Rozenblit

Sr. Manager, Manufacturing LT Israel

Issued on 30-Jan-2025