



Certificate of Analysis

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999
Sparks MD 21152-0999 US

Page: 1 of 3

Product Name : Pail Trypticase Soy Broth 25Lb
Catalog Number : 211772 **Manufacture Date**: 2024/05/30
Batch Number : 4240898
Expiration Date : 2028/04/30

01. Dehydrated Medium Powder Appearance: Fine, homogeneous, free of extraneous material
02. Solubility (30 g/L): Soluble in distilled or deionized water
03. Prepared Medium Broth Appearance: Light, tan to yellow; clear to slightly hazy
04. Biological Performance: Medium was prepared per label directions. Media is tested per European (EP), Japanese (JP) and United States Pharmacopeia (USP) Growth Promotion Requirements. Tubes were inoculated with < 100 CFUs. Tubes were incubated aerobically for 3 days and up to 5 days for the (*) organisms, as indicated below, and gave cultural responses as indicated:
- | CULTURE | ATCC® | INCUBATION | RESULT |
|---------------------------|-------|-----------------|--------|
| *Aspergillus brasiliensis | 16404 | 20-25°C | Growth |
| Bacillus subtilis | 6633 | 20-25°C/30-35°C | Growth |
| *Candida albicans | 10231 | 20-25°C | Growth |
| Escherichia coli | 8739 | 30-35°C | Growth |
| Pseudomonas aeruginosa | 9027 | 30-35°C | Growth |
| Salmonella typhimurium | 14028 | 30-35°C | Growth |
| Staphylococcus aureus | 6538 | 30-35°C | Growth |
05. Residual Solvents (CPMP/ICH/283/95): Typical Analysis and supplier-provided information for Trypticase Soy Broth indicates that there is less than 5000 ppm of Acetone, and less than 890 ppm of Toluene. No other solvents were detected during analysis.

Characteristic	Unit	Value	Lower Limit	Upper Limit
Loss on Drying :	%	2		5
pH at 25°C :		7.4	7.1	7.5
Bulk Lot Number :	-	4141979		

Animal source	Country of Origin	Tissue Category		
		BIC	SIC	ABC
Bovine	New Zealand	IV	IV	MLK
Porcine	USA	III	III	IB

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostics - Diagnostic Systems products are manufactured in ISO 13485:2016 Registered facilities. In addition, BD Diagnostics -

Creation Date: 2024/09/18 16:32:27



Certificate of Analysis

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999
Sparks MD 21152-0999 US

Page: 2 of 3

Product Name : Pail Trypticase Soy Broth 25Lb
Catalog Number : 211772 **Manufacture Date**: 2024/05/30
Batch Number : 4240898
Expiration Date : 2028/04/30

Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at www.bd.com/regdocs. For each batch of finished product that contains animal origin raw materials, the COA shows the animal origin data from the individual lots of animal origin raw materials used, as provided by the raw material suppliers.

At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at www.bd.com/regdocs under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program.

For complete details on animal origin information, refer to "BD Diagnostics - Diagnostic Systems, Animal Origin Position Statement", at www.bd.com/regdocs under "Position Statements".

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see www.bd.com/en-us/support/bd-life-sciences-diagnostic-systems-customer-regulatory-support-information.

Stability data for this product is on file.
The pH stated was obtained at the time of release, and the pH may vary within the stated range depending on the age of the product and the type of pH meter and probe used.



Certificate of Analysis

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999
Sparks MD 21152-0999 US

Page: 3 of 3

Product Name	: Pail Trypticase Soy Broth 25Lb	
Catalog Number	: 211772	Manufacture Date: 2024/05/30
Batch Number	: 4240898	
Expiration Date	: 2028/04/30	

A handwritten signature in black ink, appearing to be "Stacie Wright", written in a cursive style.

Stacie Wright
Associate Director, Quality Assurance
Signature Date: 2024/09/18