

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the purchaser, formulator or those performing further manufacturing to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalogue Number:	34843	Quality Test / Release Date:	22-May-24
Lot Number:	2497400	Expiry Phrase:	Retest after 2 years if unopened
Product of:	Saudi Arabia		

Description:	Dimethylformamide extra dry		
Grade:		Application:	

Appearance:	Clear colourless Mobile liquid
--------------------	--------------------------------

Result Name	Test Value	Units	Specification
Acidity (as CH ₃ COOH)	0.0002	%	<= 0.01
Assay (GC)	99.92	%	>= 99.75
Identity FTIR (Pass/Fail)	Pass		Must be 'Pass'
Water	0.0008	%	<= 0.005

Additional Information:	
--------------------------------	--



Paula Faulkes

Mrs Paula Faulkes
Quality Assurance Manager

Monday, 23 December, 2024

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above. If there are any questions with this certificate, please call Customer Services on +44(0)1509 555500